

AQUASOL DIGITAL

WaterProof Hand Held ORP Meter



An
ISO 9001:2015
Company



AQUASOL DIGITAL

WaterProof Hand Held ORP Meter

Kindly inspect the equipment for any signs of damage and verify that you have received all the equipments

Procedure for Calibration

Calibration is not required

Note

1. Meter can be tested with a known ORP solution to check functionality of the electrode.
2. Change a new battery when the display fades or flashes.

Procedure for Measurement

1. Rinse the electrode with clean water and wipe it dry. Dip the electrode into sample solution to be measured. Stir gently and wait until a stable reading can be obtained.
2. After measurement, rinse the electrode with clean water, and replace the protective cap.

WaterProof Hand Held ORP Meter

Procedure for Maintenance

Battery replacement

1. Loosen the battery compartment counterclockwise.
2. Remove the old Lithium battery CR2032 and note polarity.
3. Replace the battery according to the polarity & close the battery compartment cap tightly.

Electrode replacement

1. Unscrew the electrode collar clockwise, and remove it completely.
2. Pull the electrode module out from the tester.
3. Plug in new electrode module into the tester socket carefully.
4. Replace and tighten the electrode collar to make a good seal.



WaterProof Hand Held
ORP Meter

Warranty Card*

* Warranty For Meter Electronic Module Only

Customer Name/ Address : _____

Dealer Name/ Address: _____

Bill No. _____

Date -: _____



Product details



RAKIRO BIOTECH SYSTEMS PVT LTD

Web : www.rakiro.net